
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
BabyBio™ Q 5 ml	45 100 105
	45 100 106
Lot number: 155662	45 100 107
	45 100 108
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity*:	191 µmol Cl ⁻ /ml	180-250 µmol Cl ⁻ /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-06-07

Expiry date (year-month): 2027-02

Manufacturing date (year-month): 2022-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-06-07

By Anna Boman



Quality Assurance
