
CERTIFICATE OF ANALYSIS

Product:	BabyBio™ Co-NTA 1 ml	Article number(s):	45 655 131 45 655 132 45 655 133 45 655 134
Lot number:	155733		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Co ²⁺ ∗:	25.5 µmol Cu ²⁺ /ml	20-30 µmol Cu ²⁺ /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-05-10

Expiry date (year-month): 2027-04

Manufacturing date (year-month): 2022-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2022-05-10

By Anna Boman



Quality Assurance
