
CERTIFICATE OF ANALYSIS

Product:	BabyBio™ DEAE 5 ml	Article number(s):	45 150 105 45 150 106 45 150 107 45 150 108
Lot number:	155766		
Storage solution:			20% ethanol

	Measured Value	Specification
Ionic capacity*:	127 µmol Cl ⁻ /ml	110-160 µmol Cl ⁻ /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-05-23

Expiry date (year-month): 2027-04

Manufacturing date (year-month): 2022-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-05-23

By Susann Sandström

DocuSigned by:

Susann Sandstrom

8DD857460900400...

Quality Assurance
