

---

## CERTIFICATE OF ANALYSIS

---

Product:	BabyBio™ S 5 ml	Article number(s):	45 200 105 45 200 106 45 200 107 45 200 108
Lot number:	155767		
Storage solution:	20% ethanol, 0.2 M sodium acetate		

---

	Measured Value	Specification
Ionic capacity*:	233 µmol H <sup>+</sup> /ml	180-250 µmol H <sup>+</sup> /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2022-05-23

Expiry date (year-month): 2026-05


Manufacturing date (year-month): 2021-04

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-05-23

By Susann Sandström

DocuSigned by:  
  
8DDB57460000400...

Quality Assurance

---