
CERTIFICATE OF ANALYSIS

Product:	BabyBio™ S 5 ml	Article number(s):	45 200 105 45 200 106 45 200 107 45 200 108
Lot number:	155767		
Storage solution:	20% ethanol, 0.2 M sodium acetate		

	Measured Value	Specification
Ionic capacity*:	233 µmol H ⁺ /ml	180-250 µmol H ⁺ /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-05-23	Expiry date (year-month): 2026-05
	Manufacturing date (year-month): 2021-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.


obsolete due to name change CC-22-10384


Issued (year-month-day): 2022-05-23

Philip Eriksson

By Susann Sandström

2022-08-31

DocuSigned by:

8DDB57460000400...

DocuSigned by:

158C686C7E5E4A0...

Quality Assurance
