
CERTIFICATE OF ANALYSIS

Product: GoBio™ Mini S 5 mL
Lot number: 155767
Storage solution: 20% ethanol, 0.2 M sodium acetate

Article number(s):
45 200 105
45 200 106
45 200 107
45 200 108

	Measured Value	Specification
Ionic capacity*:	233 µmol H ⁺ /mL	180-250 µmol H ⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-05-23

Expiry date (year-month): 2026-05


Manufacturing date (year-month): 2021-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-09-01

By Philip Eriksson

DocuSigned by:



158C686C7E5E4A0...

Quality Assurance
