
CERTIFICATE OF ANALYSIS

Product:	BabyBio™ Ni-NTA 1 ml	Article number(s):	45 655 101 45 655 102 45 655 103 45 655 104 45 655 110
Lot number:	155800		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ²⁺ ∗:	25 µmol Cu ²⁺ /ml	20-30 µmol Cu ²⁺ /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-06-07

Expiry date (year-month): 2025-09

Manufacturing date (year-month): 2022-06

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2022-06-07

By Anna Boman

Quality Assurance
