

---

## CERTIFICATE OF ANALYSIS

---

Product:	BabyBio™ Q 5 ml	Article number(s):	45 100 105 45 100 106 45 100 107 45 100 108
Lot number:	155817		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Ionic capacity*:	210 µmol Cl <sup>-</sup> /ml	180-250 µmol Cl <sup>-</sup> /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2022-06-07

Expiry date (year-month): 2027-02

Manufacturing date (year-month): 2022-05

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-06-07

By [Name]



Quality Assurance

---