

---

## CERTIFICATE OF ANALYSIS

---

Product:	WorkBeads™ 40 DEAE	Article number(s):	40 150 001 40 150 002 40 150 010 40 150 050 40 150 060
Lot number:	155851		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Ionic capacity:	124 µmol Cl/ml	110-160 µmol Cl/ml
Average particle size, Dv50:	50 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

\*Property measured on base matrix.

---

Release date (year-month-day): 2022-06-14

Expiry date (year-month): 2027-06

Manufacturing date (year-month): 2022-06


---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-06-14

By Philip Eriksson

DocuSigned by:



Quality Assurance

---