
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
BabyBio™ A 5 ml	45 605 105
	45 605 106
Lot number: 155857	45 605 107
	45 605 108
Storage solution:	20% ethanol

	Measured Value	Specification
Dynamic binding capacity for IgG*:	51	>40 mg/ml
Gap control:	Passed	No gaps
Maximum flow test:	Passed	Stable pressure
Visual control:	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-06-16

Expiry date (year-month): 2025-03

Manufacturing date (year-month): 2022-06

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.

Country of origin is Sweden.

Issued (year-month-day): 2022-06-16

By Philip Eriksson

DocuSigned by:

158C686C7E5E4A0...
Quality Assurance
