

---

## CERTIFICATE OF ANALYSIS

---

Product:	WorkBeads™ affimAb	Article number(s):	40 800 001 40 800 002 40 800 010 40 800 050 40 800 060
Lot number:	155883		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Dynamic binding capacity for IgG:	44 mg/ml	>40 mg/ml
Average particle size, Dv50:	45 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

\*Property measured on base matrix.

---

Release date (year-month-day): 2022-06-28

Expiry date (year-month): 2025-06


Manufacturing date (year-month): 2022-06

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-06-28

By Philip Eriksson

DocuSigned by:  
  
158C686C7E5E4A0...  
Quality Assurance

---