
CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 200 SEC SG** Article number(s):
20 300 210
20 300 250

Lot number: 155901

Storage solution: 0.1 M sodium hydroxide

	Measured Value	Specification
K _D (Ovalbumin):	0,7	0.62-0.75
K _D (Thyroglobolin):	0,29	0.25-0.45
Average particle size, Dv50:	187 µm	170-210 µm
Particles within range 130-250 µm*	93 % (vol)	≥85 % (vol)

*Property measured on base matrix.

Release date (year-month-day): 2022-06-23

Expiry date (year-month): 2025-06


Manufacturing date (year-month): 2022-06

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-06-23

By Philip Eriksson

DocuSigned by:



158C686C7E5E4A0...
Quality Assurance
