
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40 Q	40 100 001
	40 100 002
	40 100 010
Lot number: 155915	40 100 050
	40 100 060
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity:	201 µmol Cl/ml	180-250 µmol Cl/ml
Average particle size, Dv50:	53 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2022-07-01


Expiry date (year-month): 2027-06

Manufacturing date (year-month): 2022-06

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-07-01

By Philip Eriksson

DocuSigned by:

158C686C7E5E4A0...
Quality Assurance
