

---

## CERTIFICATE OF ANALYSIS

---

Product:	Article number(s):
GoBio™ Mini A 5 mL	45 605 105
	45 605 106
Lot number: 155935	45 605 107
	45 605 108
Storage solution:	20% ethanol

---

	Measured Value	Specification
Dynamic binding capacity for IgG*:	49.2 mg/mL	>40 mg/mL
Gap control:	Passed	No gaps
Maximum flow test:	Passed	Stable pressure
Visual control:	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2022-08-23

Expiry date (year-month): 2023-09

Manufacturing date (year-month): 2020-09

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.

Country of origin is Sweden.

Issued (year-month-day): 2022-08-23

By Anna Boman

DocuSigned by:



3C2F4EE14561428...

Quality Assurance

---