
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
OptioBio™40Q 10x100	55 410 011
Lot number: 155986	
Column ID:	02
Resin lot number:	155786
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity*:	192 µmol Cl/ml	180-250 µmol Cl/ml
Efficiency:	6950 N/m	>6350 N/m
Asymmetry:	0,85	0.75-1.50
Bed height:	102 mm	100-102 mm
Column backpressure: (at max recommended flow)	0,06 MPa	Not specified

*Property measured on resin.

Release date (year-month-day): 2022-08-22	Expiry date (year-month): 2027-05
	Manufacturing date (year-month): 2022-07

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.

Country of origin is Sweden.

Issued (year-month-day): 2022-08-22

By Anna Boman

DocuSigned by:

3C2E4FE14561428

Quality Assurance
