

---

## CERTIFICATE OF ANALYSIS

---

Product:	BabyBio™ Q 1 ml	Article number(s):	45 100 101 45 100 102 45 100 103 45 100 104
Lot number:	155992		
Storage solution:		20% ethanol	

---

	Measured Value	Specification
Ionic capacity*:	201 µmol Cl <sup>-</sup> /ml	180-250 µmol Cl <sup>-</sup> /ml
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2022-08-22

Expiry date (year-month): 2027-06

Manufacturing date (year-month): 2022-06

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-08-25

By Anna Boman

DocuSigned by:



3C2F4EE14561428

---

Quality Assurance

---