

---

## CERTIFICATE OF ANALYSIS

---

Product:	GoBio™ Mini affimAb 1 mL	Article number(s):	45 800 101 45 800 102 45 800 103 45 800 104
Lot number:	156092		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Dynamic binding capacity for IgG*:	44 mg/mL	>40 mg/mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2022-09-02

Expiry date (year-month): 2025-06

Manufacturing date (year-month): 2022-08

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-09-02

By Philip Eriksson

DocuSigned by:



---

158C686C7E5E4A0...  
Quality Assurance