
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini DEAE 1 mL	Article number(s):	45 150 101 45 150 102 45 150 103 45 150 104
Lot number:	156137		
Storage solution:			20% ethanol

	Measured Value	Specification
Ionic capacity*:	127 µmol Cl ⁻ /mL	110-160 µmol Cl ⁻ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2022-09-26


Expiry date (year-month): 2027-04

Manufacturing date (year-month): 2022-09

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-09-26

By Philip Eriksson

DocuSigned by:

158C686C7E5E4A0...
Quality Assurance
