
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40 S	40 200 001
	40 200 002
	40 200 005
Lot number: 156142	40 200 010
	40 200 050
	40 200 060
Storage solution:	20% ethanol, 0.2M sodium acetate

	Measured Value	Specification
Ionic capacity:	237 µmol H ⁺ /ml	180-250 µmol H ⁺ /ml
Average particle size, Dv50:	51 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2022-09-15

Expiry date (year-month): 2027-09

Manufacturing date (year-month): 2022-09

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-09-16

By Philip Eriksson

DocuSigned by:



158C686C7E5E4A0...

Quality Assurance
