
CERTIFICATE OF ANALYSIS

| | |
|--------------------|--------------------|
| Product: | Article number(s): |
| WorkBeads™ 40 Q | 40 100 001 |
| | 40 100 002 |
| | 40 100 010 |
| Lot number: 156205 | 40 100 050 |
| | 40 100 060 |
| Storage solution: | 20% ethanol |

| | Measured Value | Specification |
|--|------------------------------|----------------------------------|
| Ionic capacity: | 200 µmol Cl ⁻ /ml | 180-250 µmol Cl ⁻ /ml |
| Average particle size, Dv50: | 53 µm | 32-60 µm |
| Flow stability: (linear flow rate in a 10 x 300 mm column)* | >10 cm/min | ≥10 cm/min |

*Property measured on base matrix.

Release date (year-month-day): 2022-10-11


Expiry date (year-month): 2027-10

Manufacturing date (year-month): 2022-10

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-10-11

By Philip Eriksson

DocuSigned by:

158C686C7E5E4A0...

Quality Assurance
