

---

## CERTIFICATE OF ANALYSIS

---

Product:	WorkBeads™ 40/1000 SEC	Article number(s):	40 300 001 40 300 003 40 300 010 40 300 050 40 300 060
Lot number:	156214		
Storage solution:			20% ethanol

---

	Measured Value	Specification
K <sub>D</sub> (Ovalbumin):	0,64	0.62-0.75
Average particle size, Dv50:	51 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

\*Property measured on base matrix.

---

Release date (year-month-day): 2022-10-27

Expiry date (year-month): 2027-10

Manufacturing date (year-month): 2022-10

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-11-09

By Philip Eriksson

DocuSigned by:



158C886C7E5E4A0...

Quality Assurance

---