
CERTIFICATE OF ANALYSIS

| | | | |
|-------------------|-----------------------|--------------------|--|
| Product: | GoBio™ Mini TREN 1 mL | Article number(s): | 45 655 211 45 655 212 45 655 213 45 655 214 |
| Lot number: | 156242 | | |
| Storage solution: | | 20% ethanol | |

| | Measured Value | Specification |
|-------------------|------------------------------|----------------------------------|
| Ionic capacity*: | 163 µmol Cl ⁻ /mL | 130-200 µmol Cl ⁻ /mL |
| Gap control: | Passed | No gaps |
| Maximum flow test | Passed | Stable pressure |
| Visual control | Passed | No visible defects |

*Property measured on resin.

Release date (year-month-day): 2023-03-21

Expiry date (year-month): 2025-04

Manufacturing date (year-month): 2022-10

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-03-21

By Jonatan Bylund

DocuSigned by:



F0489A1BCFA8414...

Quality Assurance
