
CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 SEC Cube** Article number(s): 40 301 050

Lot number: 156304

Storage solution: 20% ethanol

	Measured Value	Specification
K _D (Thyroglobulin):	0,29	0.20-0.30
Average particle size, Dv50:	50 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2022-11-16

Expiry date (year-month): 2027-11

Manufacturing date (year-month): 2022-11

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-01-09

By Jonatan Bylund

DocuSigned by:



F0489A1BCFA8414...

Quality Assurance
