
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 100 Q	10 100 001
	10 100 002
	10 100 005
Lot number: 156392	10 100 010
	10 100 050
	10 100 060
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity:	160 µmol Cl/ml	140-200 µmol Cl/ml
Average particle size, Dv50:	93 µm	90-110 µm
Flow stability: (linear flow rate in a 25 x 200 mm column)*	1200 cm/hr	≥1000 cm/hr

*Property measured on base matrix.

Release date (year-month-day): 2022-12-20

Expiry date (year-month): 2027-12

Manufacturing date (year-month): 2022-12

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-12-20

By Jonatan Bylund

DocuSigned by:



F0489A1BCFA8414...

Quality Assurance
