
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Ni-NTA 5 mL	Article number(s):	45 655 105 45 655 106 45 655 107 45 655 108 45 655 109
Lot number:	156450		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ^{2+*} :	25 µmol Cu ²⁺ /mL	20-30 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2023-01-31

Expiry date (year-month): 2027-02

Manufacturing date (year-month): 2022-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-01-31

By Jonatan Bylund

DocuSigned by:



F0489A1BCFA8414...

Quality Assurance
