



CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40 DEAE	40 150 001
	40 150 002
	40 150 010
Lot number: 156469	40 150 050
	40 150 060
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity:	147 µmol Cl ⁻ /ml	110-160 µmol Cl ⁻ /ml
Average particle size, Dv50:	52 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2023-02-08

Expiry date (year-month): 2028-02

Manufacturing date (year-month): 2023-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-02-14

By Jonatan Bylund

DocuSigned by:

F0489A1BCFA8414...

Quality Assurance