
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini S 1 mL	Article number(s):	45 200 101 45 200 102 45 200 103 45 200 104
Lot number:	156596		
Storage solution:	20% ethanol, 0.2 M sodium acetate		

	Measured Value	Specification
Ionic capacity*:	202 µmol H ⁺ /mL	180-250 µmol H ⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2023-04-06


Expiry date (year-month): 2024-08

Manufacturing date (year-month): 2023-03

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-04-06

By Philip Eriksson

DocuSigned by:

158C686C7E5E4A0...
Quality Assurance
