
CERTIFICATE OF ANALYSIS

Product:	WorkBeads™ 40/1000 SEC	Article number(s):	40 300 001 40 300 003 40 300 010 40 300 050 40 300 060
Lot number:	156679		
Storage solution:			20% ethanol

	Measured Value	Specification
K _D (Ovalbumin):	0.62	0.62-0.75
Average particle size, Dv50:	53 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2023-09-08


Expiry date (year-month): 2028-07

Manufacturing date (year-month): 2023-07

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-10-05

By Jonatan Bylund

DocuSigned by:

F0489A1BCFA8414...
Quality Assurance
