
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Q 5 mL	Article number(s):	45 100 105 45 100 106 45 100 107 45 100 108
Lot number:	156716		
Storage solution:			20% ethanol

	Measured Value	Specification
Ionic capacity*:	197 µmol Cl ⁻ /mL	180-250 µmol Cl ⁻ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2023-05-22

Expiry date (year-month): 2028-04

Manufacturing date (year-month): 2023-05

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-05-22

By Jonatan Bylund

DocuSigned by:



F0489A1BCFA8414...

Quality Assurance
