

---

## CERTIFICATE OF ANALYSIS

---

Product:	GoBio™ Mini NiMAC 1 mL	Article number(s):	45 655 311 45 655 312 45 655 313 45 655 314
Lot number:	156722		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Metal binding capacity before charging with Ni <sup>2+*</sup> :	115 µmol Cu <sup>2+</sup> /mL	>60 µmol Cu <sup>2+</sup> /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2023-05-16

Expiry date (year-month): 2027-05

Manufacturing date (year-month): 2023-05

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2023-05-16

By Philip Eriksson

DocuSigned by:



158C686C7E5E4A0...

Quality Assurance

---