

---

## CERTIFICATE OF ANALYSIS

---

Product:	GoBio™ Mini S 1 mL	Article number(s):	45 200 101 45 200 102 45 200 103 45 200 104
Lot number:	156724		
Storage solution:	20% ethanol, 0.2 M sodium acetate		

---

	Measured Value	Specification
Ionic capacity*:	250 µmol H <sup>+</sup> /mL	180-250 µmol H <sup>+</sup> /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2023-05-22

Expiry date (year-month): 2028-04

Manufacturing date (year-month): 2023-05


---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2023-05-22

By Philip Eriksson

DocuSigned by:



158C686C7E5E4A0...

Quality Assurance

---