

---

## CERTIFICATE OF ANALYSIS

---

Product:	GoBio™ Mini Q 5 mL	Article number(s):	45 100 105 45 100 106 45 100 107 45 100 108
Lot number:	156769		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Ionic capacity*:	198 µmol Cl <sup>-</sup> /mL	180-250 µmol Cl <sup>-</sup> /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2023-06-08

Expiry date (year-month): 2028-05

Manufacturing date (year-month): 2023-05


---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2023-06-08

By Philip Eriksson

DocuSigned by:



Quality Assurance

---