

---

## CERTIFICATE OF ANALYSIS

---

|                   |                    |                    |  |
|-------------------|--------------------|--------------------|--|
| Product:          | GoBio™ Mini Q 5 mL | Article number(s): | 45 100 105<br>45 100 106<br>45 100 107<br>45 100 108 |
| Lot number:       | 156837             |                    |  |
| Storage solution: |                    |                    | 20% ethanol  |

---

|                   | Measured Value               | Specification                    |
|-------------------|------------------------------|----------------------------------|
| Ionic capacity*:  | 198 µmol Cl <sup>-</sup> /mL | 180-250 µmol Cl <sup>-</sup> /mL |
| Gap control:      | Passed                       | No gaps                          |
| Maximum flow test | Passed                       | Stable pressure                  |
| Visual control    | Passed                       | No visible defects               |

\*Property measured on resin.

---

Release date (year-month-day): 2023-06-15

Expiry date (year-month): 2028-05

Manufacturing date (year-month): 2023-06


---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2023-06-15

By Philip Eriksson

DocuSigned by:



158C686C7E5E4A0...

Quality Assurance

---