
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini NiMAC 5 mL	Article number(s):	45 655 315 45 655 316 45 655 317 45 655 318
Lot number:	156856		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ²⁺ ∗:	115 µmol Cu ²⁺ /mL	>60 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2023-06-21

Expiry date (year-month): 2027-05

Manufacturing date (year-month): 2023-06

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2023-10-26

By Jonatan Bylund

DocuSigned by:

F0489A1BCFA8414...
Quality Assurance
