
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini NiMAC 5 mL	Article number(s):	45 655 315 45 655 316 45 655 317 45 655 318
Lot number:	156856		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ^{2+*} :	115 µmol Cu ²⁺ /mL	>60 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2023-06-21

Expiry date (year-month): 2027-05


Manufacturing date (year-month): 2022-06


This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2023-06-21

By Philip Eriksson

This CoA is obsolete, the manufacturing date was entered incorrectly, see NC-23-10571, related to complaint 10571.

DocuSigned by:

158C688C7E5E4A0...
Quality Assurance

DocuSigned by:

2023-10-26
F0489A1BCEA8414...