
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
GoBio™ Mini S 5 mL	45 200 105
	45 200 106
Lot number: 156901	45 200 107
	45 200 108
Storage solution:	20% ethanol, 0.2 M sodium acetate

	Measured Value	Specification
Ionic capacity*:	250 µmol H ⁺ /mL	180-250 µmol H ⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2023-07-17

Expiry date (year-month): 2028-04

Manufacturing date (year-month): 2023-07

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-08-08

By Jonatan Bylund

DocuSigned by:



F0489A1BCFA8414...

Quality Assurance
