
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
GoBio™ Mini S 1 mL	45 200 101
	45 200 102
Lot number: 156999	45 200 103
	45 200 104
Storage solution:	20% ethanol, 0.2 M sodium acetate

	Measured Value	Specification
Ionic capacity*:	250 µmol H ⁺ /mL	180-250 µmol H ⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2023-10-13

Expiry date (year-month): 2028-09

Manufacturing date (year-month): 2023-09

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-10-13

By Jonatan Bylund

DocuSigned by:



F0489A1BCFA8414...

Quality Assurance
