

---

## CERTIFICATE OF ANALYSIS

---

Product: GoBio™ Mini NTA His-Tag Screening Kit 1 mL  
Article number(s): 45 700 101

Lot number: 157006

Kit consists of: Lot number:  
GoBio Mini Ni-NTA 1 mL: 155800  
GoBio Mini Cu-NTA 1 mL: 154122  
GoBio Mini Co-NTA 1 mL: 154137  
GoBio Mini Zn-NTA 1 mL: 154172

Storage solution: 20% ethanol

---

	Measured Value	Specification
Metal binding capacity before binding with metal ion:		
GoBio Mini Ni-NTA*:	25 µmol/mL	20-30 µmol Cu <sup>2+</sup> /mL
GoBio Mini Cu-NTA*:	25 µmol/mL	20-30 µmol Cu <sup>2+</sup> /mL
GoBio Mini Co-NTA*:	25 µmol/mL	20-30 µmol Cu <sup>2+</sup> /mL
GoBio Mini Zn-NTA*:	25 µmol/mL	20-30 µmol Cu <sup>2+</sup> /mL
Gap control:	Passed	No gaps
Maximum flow test:	Passed	Stable pressure
Visual control:	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2020-10-01

Expiry date (year-month): 2025-05

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2023-10-03

By Jonatan Bylund

DocuSigned by:



FD489A1BCFA8414...

Quality Assurance

---