
CERTIFICATE OF ANALYSIS

Product:	WorkBeads™ 40 DEAE	Article number(s):	40 150 001 40 150 002 40 150 010 40 150 050 40 150 060
Lot number:	157012		
Storage solution:			20% ethanol

	Measured Value	Specification
Ionic capacity:	136 µmol Cl/ml	110-160 µmol Cl/ml
Average particle size, Dv50:	49 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	>10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2023-10-24

Expiry date (year-month): 2028-10

Manufacturing date (year-month): 2023-10

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-10-24

By Jonatan Bylund

DocuSigned by:



F0489A1BCFA8414...

Quality Assurance
