

---

## CERTIFICATE OF ANALYSIS

---

|                   |                          |                    |  |
|-------------------|--------------------------|--------------------|--|
| Product:          | GoBio™ Mini affimAb 5 mL | Article number(s): | 45 800 105<br>45 800 106<br>45 800 107<br>45 800 108 |
| Lot number:       | 157131                   |                    |  |
| Storage solution: |                          |                    | 20% ethanol  |

---

|                                    | Measured Value | Specification      |
|------------------------------------|----------------|--------------------|
| Dynamic binding capacity for IgG*: | 44.4 mg/mL     | >40.0 mg/mL        |
| Gap control:                       | Passed         | No gaps            |
| Maximum flow test                  | Passed         | Stable pressure    |
| Visual control                     | Passed         | No visible defects |

\*Property measured on resin.

---

Release date (year-month-day): 2023-11-22

Expiry date (year-month): 2026-09

Manufacturing date (year-month): 2023-11

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2023-11-22

By Åsa Nilsson

DocuSigned by:

Åsa Nilsson

ED9074A2ACBC431

---

Quality Assurance

---