



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini S 5 mL	Article number(s):	45 200 105
			45 200 106
			45 200 107
			45 200 108
Lot number: 157157			
Storage solution:	20% ethanol, 0.2 M sodium acetate		

	Measured Value	Specification
Ionic capacity*:	250 µmol H ⁺ /mL	180-250 µmol H ⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2023-12-13

Expiry date (year-month): 2028-11

Manufacturing date (year-month): 2023-11

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2023-12-13

By [Name]

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
