



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Ni-NTA 1 mL	Article number(s):	45 655 101 45 655 102 45 655 103 45 655 104 45 655 110
Lot number:	157159		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ²⁺ ∗:	25 µmol Cu ²⁺ /mL	20-30 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2024-01-16

Expiry date (year-month): 2028-03

Manufacturing date (year-month): 2023-03

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2024-01-16

By [Name]

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance