
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini ACT 1 mL	Article number(s):	45 400 001 45 400 002 45 400 003 45 400 004
Lot number:	157216		
Storage solution:			20% ethanol

	Measured Value	Specification
Concentration of reactive groups*:	256 µmol/mL	>200 µmol/mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2024-01-16

Expiry date (year-month): 2025-03

Manufacturing date (year-month): 2024-01

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-01-16

By Åsa Nilsson

DocuSigned by:

FD9074A2ACBC431...

Quality Assurance
