



---

## CERTIFICATE OF ANALYSIS

---

Product:  
GoBio™ Prep 16x600 Macro SEC

Article number:  
55 437 026

Lot number:D08012402

Storage solution:

20% ethanol

---

	Measured Value	Specification
K <sub>D</sub> (Ovalbumin)*:	0.78	0.70-0.90
Efficiency:	7242 N/m	≥5000 N/m
Asymmetry:	0.83	0.7-1.4
Bed height:	605 mm	600-620 mm

\*Property measured on resin.

Resin lot number: 156432

---

Release date (year-month-day): 2024-01-15

Expiry date (year-month): 2028-01

Manufacturing date (year-month): 2024-01

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2024-01-15

By [Name]

DocuSigned by:

*Markus Norén*

6853712A40184AF...

Quality Assurance