



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Butyl SH 5 mL	Article number(s):	45 500 105 45 500 107 45 500 108
Lot number:	157248		
Storage solution:			20% ethanol

	Measured Value	Specification
Ligand density*:	55 µmol/mL	46-62 µmol/mL
Gap control:	Passed	No gaps
Maximum flow test:	Passed	Stable pressure
Visual control:	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2024-02-15

Expiry date (year-month): 2028-12

Manufacturing date (year-month): 2023-11

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-02-15

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
