
CERTIFICATE OF ANALYSIS

Product: **GoBio™ Mini IEX Screening Kit** Article number(s): 45 900 001

Lot number: 157326

Kit consists of:	Lot number:
GoBio Mini S 1 mL:	156999
GoBio Mini Q 1 mL:	156964
GoBio Mini DEAE 1 mL:	157068
GoBio Mini TREN 1 mL:	156242

Storage solution: 20% ethanol (0.2 M sodium acetate added to GoBio Mini S)

	Measured Value	Specification
Ionic capacity (GoBio Mini S)*:	250 µmol H ⁺ /mL	180-250 µmol H ⁺ /mL
Ionic capacity (GoBio Mini Q)*:	198 µmol Cl ⁻ /mL	180-250 µmol Cl ⁻ /mL
Ionic capacity (GoBio Mini DEAE)*:	127 µmol Cl ⁻ /mL	110-160 µmol Cl ⁻ /mL
Ionic capacity (GoBio Mini TREN)*:	163 µmol Cl ⁻ /mL	130-200 µmol Cl ⁻ /mL
Gap control:	Passed	No gaps
Maximum flow test:	Passed	Stable pressure
Visual control:	Passed	No visible defects

*Property measured on resin.


Release date (year-month-day): 2024-10-17

Expiry date (year-month): 2025-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-10-17

By Åsa Nilsson

DocuSigned by:

FD9074A2ACBC431...
Quality Assurance
