



CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40S** Article number(s):
40 200 001
40 200 002
40 200 010
40 200 050
40 200 060

Lot number: 157338

Storage solution: 20% ethanol, 0.2 M sodium acetate

	Measured Value	Specification
Ionic capacity:	182 $\mu\text{mol H}^+/\text{ml}$	180-250 $\mu\text{mol H}^+/\text{ml}$
Average particle size, Dv50:	47 μm	32-60 μm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	≥ 10 cm/min	≥ 10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2024-03-12

Expiry date (year-month): 2029-02

Manufacturing date (year-month): 2024-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-03-12

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance