
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Q 1 mL	Article number(s):	45 100 101 45 100 102 45 100 103 45 100 104
Lot number:	157340		
Storage solution:			20% ethanol

	Measured Value	Specification
Ionic capacity*:	198 µmol Cl ⁻ /mL	180-250 µmol Cl ⁻ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2024-03-28

Expiry date (year-month): 2028-05

Manufacturing date (year-month): 2023-05

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-03-28

By Åsa Nilsson

DocuSigned by:

Åsa Nilsson

FD9074A2ACBC431

Quality Assurance
