

---

## CERTIFICATE OF ANALYSIS

---

Product: GoBio™ Mini S 1 mL  
Lot number: 157450  
Storage solution: 20% ethanol, 0.2 M sodium acetate

Article number(s):  
45 200 101  
45 200 102  
45 200 103  
45 200 104

---

	Measured Value	Specification
Ionic capacity*:	207 µmol H <sup>+</sup> /mL	180-250 µmol H <sup>+</sup> /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2024-05-08

Expiry date (year-month): 2029-02

Manufacturing date (year-month): 2024-03

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2024-05-14

By Åsa Nilsson

DocuSigned by:



FD9074A2ACBC431...

Quality Assurance

---