



CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40 Q	40 100 001
	40 100 002
	40 100 010
Lot number: 157485	40 100 050
	40 100 060
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity:	203 µmol Cl-/ml	180-250 µmol Cl-/ml
Average particle size, Dv50:	45 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	≥10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2024-04-17

Expiry date (year-month): 2029-04

Manufacturing date (year-month): 2024-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-04-17

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance